

# Meridian Gel-Lite Gel/Foam Overlay

SKU: GO-78343-C

Warranty: 1 year on Overlay and Cover



MERIDIAN  
MEDICAL™

## DIRECTIONS FOR USE:

Place the Gel-Lite Gel/Foam Overlay on top of the existing bed mattress with the blue side facing up. The elastic straps must be placed under the **bed mattress** to prevent the Gel-Lite Gel/Foam Overlay from shifting.

**Due to compression from shipping, the Gel-Lite Gel/Foam Overlay will need to decompress. The Gel-Lite Gel/Foam Overlay will begin to decompress to full length after being attached to the bed mattress.**

## CLEANING INSTRUCTIONS:

Wipe clean using a mild disinfectant if required. Spills or fluids should be wiped up immediately for patient hygiene. A bed sheet should be used to cover the mattress overlay. To order fitted mattress overlay sheets please ask your health care dealer.

Any questions or comments should be directed to your health care dealer.

**Maximum User Weight Capacity: 250 lbs.**



## WARNING

### ENTRAPMENT WARNING

Whenever a support surface and bed rails are used as part of any hospital or home care bed system, the potential for bed rail entrapment exists. To reduce the risk, a patient must always be assessed regarding their risk of entrapment before any product is used. Routine patient monitoring is also required once a product is in use. To learn more about the risk of bed rail entrapment, contact your Meridian Medical sales representative for a copy of our "Bed Rail

Safety Guide" or visit the U.S. FDA's website at [www.fda.gov](http://www.fda.gov) and search for "Bed Rail Entrapment".

### POLYURETHANE FOAM IS FLAMMABLE

Sources of heat and fire including smoldering cigarettes, can cause urethane foam to ignite. Urethane foam may burn rapidly, consuming oxygen at a high rate and emitting toxic gases, either of which can be incapacitating or in certain conditions, even fatal.

**1.800.376.7263 | WWW.COMPASSHEALTHBRANDS.COM**

Manufactured for Compass Health Brands

Made in China

CUT

DO NOT  
PRINT

CUT

DO NOT  
PRINT

# Meridian Gel-Lite Gel/Foam Overlay

SKU: GO-78343-C

Warranty: 1 year on Overlay and Cover



MERIDIAN  
MEDICAL™

## DIRECTIONS FOR USE:

Place the Gel-Lite Gel/Foam Overlay on top of the existing bed mattress with the blue side facing up. The elastic straps must be placed under the **bed mattress** to prevent the Gel-Lite Gel/Foam Overlay from shifting.

**Due to compression from shipping, the Gel-Lite Gel/Foam Overlay will need to decompress. The Gel-Lite Gel/Foam Overlay will begin to decompress to full length after being attached to the bed mattress.**

## CLEANING INSTRUCTIONS:

Wipe clean using a mild disinfectant if required. Spills or fluids should be wiped up immediately for patient hygiene. A bed sheet should be used to cover the mattress overlay. To order fitted mattress overlay sheets please ask your health care dealer.

Any questions or comments should be directed to your health care dealer.

**Maximum User Weight Capacity: 250 lbs.**



## WARNING

### ENTRAPMENT WARNING

Whenever a support surface and bed rails are used as part of any hospital or home care bed system, the potential for bed rail entrapment exists. To reduce the risk, a patient must always be assessed regarding their risk of entrapment before any product is used. Routine patient monitoring is also required once a product is in use. To learn more about the risk of bed rail entrapment, contact your Meridian Medical sales representative for a copy of our "Bed Rail

Safety Guide" or visit the U.S. FDA's website at [www.fda.gov](http://www.fda.gov) and search for "Bed Rail Entrapment".

### POLYURETHANE FOAM IS FLAMMABLE

Sources of heat and fire including smoldering cigarettes, can cause urethane foam to ignite. Urethane foam may burn rapidly, consuming oxygen at a high rate and emitting toxic gases, either of which can be incapacitating or in certain conditions, even fatal.

**1.800.376.7263 | WWW.COMPASSHEALTHBRANDS.COM**

Manufactured for Compass Health Brands

Made in China